

The **FTD™ SARS-CoV-2, kit/96 tests** by **Siemens Healthineers (Siemens Healthcare GmbH)** is available for procurement through Supply Division. The product is offered at the FCA-based price **EUR5.50** per test. The price does not include an extraction kit which has to be procured separately. Technical and commercial description of the products is presented in Table 1 below.

Table 1 FTD™ SARS-CoV-2 Kit

UNICEF Material Number	Product Description	Product Code	Pack Size	Price per Pack, (EUR)	FCA Lead Time, <i>indicative</i>	FCA Port of Delivery
S0004003	FTD™ SARS-CoV-2 Kit/96 tests, includes: <ul style="list-style-type: none"> ▪ 1 Vial, SCoV2 PP - Primer/probe Mix (144 µL) ▪ 1 Vial, SCoV2 PC - Positive control (150 µL) ▪ 1 Vial, Negative Control (2000 µL) ▪ 1 Vial, Internal Control (350 µL) ▪ 1 Vial, 25x RT-PCR Enzyme mix (96 µL) ▪ 2 vial RT-PCR Buffer (1200 µL) 	11416284	96 tests	528.00	2 weeks	Duisburg, Germany

Intended purpose:

FTD SARS-CoV-2 is a qualitative *in vitro* nucleic acid amplification test for the detection of severe respiratory syndrome coronavirus 2 (SARS-CoV-2) nucleic acids in nasopharyngeal and oropharyngeal swabs of patients with signs and symptoms of SARS-CoV-2 infection in conjunction with clinical and epidemiological risk factors, who are suspected of (COVID-19). This test is a real-time polymerase chain reaction (RT-PCR)-based process.

Product Reference Note:

FTD™ SARS-CoV-2 Kit/32 tests by **Siemens Healthineers**

The following materials are not provided with the kit but are required:

1. RNA extraction kit is required to perform the test:
 - 280133 NucliSENS® easyMAG®, Magnetic Silica Beads
 - 280134 NucliSENS® easyMAG®, Lysis Buffer
 - 280130 NucliSENS® easyMAG®, Extraction Buffer 1
 - 280131 NucliSENS® easyMAG®, Extraction Buffer 2
 - 280132 NucliSENS® easyMAG®, Extraction Buffer 3
 - 280135 NucliSENS® easyMAG®, Disposables
 - 04801677 VERSANT® Sample Preparation 1.0 Reagents Box 1
 - 04801685 VERSANT® Sample Preparation 1.0 Reagents Box 2

2. PCR equipment with applicable software is required to run the test:
 - Biosystems 7500 Real-Time PCR System and 7500 Fast Dx Real-Time PCR.

Storage and stability conditions:

The kit must be stored at -30 to -10°C

Shelf life:

18 months from the date of manufacture

Should you have any questions regards to ordering please contact Tatsiana Ptashnik
tpashnik@unicef.org

Should you have any technical questions regards to the test please contact Wandani Sebonego
wsebonego@unicef.org

Product Reference Note:

FTD™ SARS-CoV-2 Kit/32 tests by Siemens Healthineers